

Prestige Antibodies®

Powered by  ATLAS
ANTIBODIES

Immunohistochemistry Procedure

Product Description

The Prestige Antibodies® are subjected to a standardized test procedure using specially designed tissue microarray (TMA) slides.

Preparation Instructions

Deparaffinization

Paraffin sections of 4 µm thickness are baked overnight at 50 °C. Prior to immunostaining, deparaffinization and hydration are done in xylene and graded ethanol to distilled water. During hydration, a 5 minute blocking for endogenous peroxidase is done with 0.03% (v/v) H₂O₂ in 95% ethanol.

Wash Buffer – working wash buffer contains 0.2% TWEEN® 20.

Procedure

Standard Antigen Retrieval Method

Heat Induced Epitope Retrieval (HIER) is performed by heating the TMA slides immersed in retrieval solution, pH 6, at 125 °C for 4 minutes in a pressure boiler. After boiling is completed, slides remain in the pressure boiler and are allowed to cool down to 90 °C. The total processing time is ~45 minutes.

Alternative Antigen Retrieval Methods

1. HIER performed with retrieval buffer, pH 9.
2. Enzymatic Antigen Retrieval - Enzymatic retrieval is performed by incubation of the TMA slides with Proteinase K for 10 minutes at room temperature (RT).

Standard primary antibody dilutions

- for antibody concentrations <0.06 mg/ml, 1:25
 - for antibody concentrations >0.06 mg/ml, 1:75
 - for antibody concentrations >0.1 mg/ml, 1:150
- Note: The specified working dilutions of the antibodies are to be considered as guidelines only. Optimal dilutions must be determined by the user.

Immunohistochemical staining – Performed with Lab Vision Autostainer™ 480. Incubations are performed at room temperature. Reagents are applied at a volume of 300 µl per TMA slide.

1. Rinse in wash buffer.
 2. Incubate with primary antibody for 30 minutes.
 3. Rinse 2 times in wash buffer.
 4. Incubate with peroxidase labeled polymer conjugated to a secondary antibody for 30 minutes
 5. Rinse 2 times in wash buffer.
 6. Develop for 10 minutes using diaminobenzidine (DAB) as the substrate.
 7. Rinse 2 times in distilled water
- Note: Steps 8–13 are done in a histostaining instrument (Leica Autostainer XL).
8. Counterstain in Mayer's hematoxylin for 5 minutes.
 9. Rinse 2 times in tap water.
 10. Rinse in lithium carbonate water, diluted 1:5 from saturated solution, for 1 minute.
 11. Rinse in tap water for 5 minutes.
 12. Dehydrate in graded ethanol and xylene.
 13. Coverslipping.

Prestige Antibodies is a registered trademark of Sigma-Aldrich® Biotechnology LP and Sigma-Aldrich Co. Autostainer is a trademark of Thermo Fisher Scientific. TWEEN is a registered trademark of Croda International PLC.

KB,LPG,MAM 06/09-1

Sigma brand products are sold through Sigma-Aldrich, Inc.

Sigma-Aldrich, Inc. warrants that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product(s) for their particular use. Additional terms and conditions may apply. Please see reverse side of the invoice or packing slip.