

## ProductInformation

### HEMATOXYLIN & EOSIN

(Regressive Method-Harris Hematoxylin Template)

### LEICA ST4040

Procedure HHS

Begin	100% Alcohol	100% Alcohol	100% Alcohol	100% Alcohol	95% Alcohol	Wash	Wash	Harris Hem	Harris Hem	Harris Hem	Harris Hem	Wash	Wash	Wash
Diff Soln	Wash	Wash	Scott's Tap	Wash	Wash	Eosin Y	95% Alcohol	100% Alcohol	100% Alcohol	100% Alcohol	Xylene	Xylene	End	

\* There are a total of 27 reagent stations, not counting the five holding stations for xylene at the beginning. The holding stations are not illustrated.