

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:

ProteoMas™ Apomyoglobin MALDI-MS Standard - vial of 10 nmol, (M+H⁺) 16,952.27 Da by calculation

Product Number: A8971
Lot Number: 049K6248V
Brand: SIGMA
CAS Number: 9008-45-1
MDL Number: MFCD00198077
Quality Release Date: 07 APR 2011

Test	Specification	Result
Protein (nanomoles/vial)	8.0 - 18.0	10.9
µg/vial	135.6 - 305.1	185.0
Suitability	Suitable	Suitable
Suitable for use as a standard for MALDI Mass Spectrometry.		



Carrie Cupp, Supervisor
Quality Control
St. Louis, Missouri US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.