

3050 Spruce Street, Saint Louis, MO 63103, USA


Website: www.sigmaaldrich.comEmail USA: techserv@sial.comOutside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:

D-(+)-Glucose solution - 1 mg/mL in 0.1% benzoic acid, standard for enzymatic assay kits GAGO20, GAHK20, STA20, analytical standard

Product Number: G3285
Batch Number: SLBJ7439V
Brand: FLUKA
MDL Number: MFCD00063774
Formula: C₆H₁₂O₆
Formula Weight: 180.16 g/mol
Storage Temperature: Store at 2 - 8 °C
Quality Release Date: 15 JUL 2014



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Turbidity)	Clear	Clear
Appearance (Form)	Liquid	Liquid
Concentration	0.96 - 1.04 mg/ml	0.99 mg/ml
Glucose		
Traceable to NIST standard		

Rodney Burbach, Manager
Analytical Services
St. Louis, Missouri US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.