

## Certificate of Analysis

**Product Name:** HICROME™ ENTEROCOCCI BROTH  
*BioChemika*, for microbiology  
**Product Number:** 52441  
**Product Brand:** Fluka  
**Molecular Formula:**  
**Molecular Mass:**  
**CAS Number:**

TEST	SPECIFICATION	LOT BCBB1838V RESULTS
APPEARANCE (COLOR)	FAINTLY YELLOW TO BEIGE	FAINTLY BEIGE
APPEARANCE (FORM)	POWDER	POWDER
MICROBIOLOGY TESTS	PASS AS DESCRIBED IN DATA SHEET	PASS
SOLUBILITY (COLOR)	SLIGHTLY YELLOW TO BROWNISH- -YELLOW	SLIGHTLY YELLOW
SOLUBILITY (TURBIDITY)	CLEAR TO FAINTLY TURBID	CLEAR
PH	7.3 - 7.7, AT 25 C	7.7
SOLUBILITY (METHOD)	-	18.18.59 G/L IN WATER, 121 C FOR 15 MIN
EXPIRATION DATE	-	18/SEP/14
QC RELEASE DATE	29/OCT/09	



Edeltraud Schwärzler, Manager  
Quality Control  
Buchs, Switzerland

Sigma-Aldrich guarantees the 'Sales-Specification' values only, additional lot specific tests may be included for further information. The current 'Sales-Specifications' sheet is available on request. For further inquiries, please contact our Technical Service. Sigma-Aldrich warrants, that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice for additional terms and conditions of sale. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.