


Certificate of Analysis

Product Name: MRS Broth
for microbiology
Product Number: 69966
Batch Number: BCCD3449
Brand: Sigma-Aldrich
CAS Number:
Formula:
Formula Weight:
Expiration Date: MAY 2025
Quality Release Date: 30 JUN 2020

TEST	SPECIFICATION	RESULT
APPEARANCE (COLOR)	FAINT YELLOW TO LIGHT YELLOW AND FAINT BEIGE TO LIGHT BEIGE	FAINT BEIGE
APPEARANCE (FORM)	POWDER	POWDER
SOLUBILITY (COLOR)	YELLOW AND BROWN-YELLOW AND BROWN	BROWN-YELLOW
SOLUBILITY (TURBIDITY)	CLEAR	CLEAR
PH	6.0 - 6.4, AT 25 C	6.0
SOLUBILITY (METHOD)	51.0 G/L IN WATER, 121 C FOR 15 MIN.	51.0 G/L IN WATER, 121 C FOR 15 MIN.



Dr. Reinhold Schwenninger
Quality Assurance
Buchs, Switzerland

Sigma-Aldrich warrants that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.